Phone: +91-8860605265 Email: info@astrascientific.com

Product Name: Product Code:
K-Slump Tester ASTConcrete-0036



## **Description:**

K-Slump Tester

## **Technical Specification:**

## Specification:

This device is used to determine the workability and degree of compaction of fresh concrete after being placed in the forms. It can be used for in-situ measurements or inside test moulds and forms. Results can be correlated against the slump test.

The operation is very simple: insert the tester into the concrete up to the level of the disc, after 60 seconds, a measuring rod in lowered onto the surface of the concrete and the K-slump is read directly on a scale. The calibrated hollow tube has a dia. of 20 mm.

Total length: 300 mm

Weight approx.: 500 g

K-Slump Tester for in-place measurements indicates correlation to the slump test. The probe determines workability of concrete and the degree

Phone: +91-8860605265

Email: info@astrascientific.com

of compaction. Includes correlation chart and instructions.



## **Astra Scientific**

www.astrascientific.com, Email: info@astrascientific.com

Address: K-88, 20th Street, Annanagar, Chennai, India – 600040 Phone: +91-8860605265